

## G7 Multihole Total Conventional Hip Investigation

Note: This analysis compares the G7 Multihole acetabular prosthesis with all other total conventional hip prostheses.

This prosthesis has been identified as having a significantly higher rate of revision. For a detailed explanation of the process used by the Registry that results in identification of prostheses that have a higher than anticipated rate of revision please refer to the Prostheses with Higher than Anticipated Rates of Revision chapter of the most recent AOANJRR Annual Report, <https://aoanjrr.sahmri.com/annual-reports-2024>.

Note: Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator. Procedures using prostheses with no recorded use in 2023 are excluded from the comparator.

### TABLE 1

#### Revision Rate of Primary Total Conventional Hip Replacement

The revision rate of the G7 Multihole total conventional hip prosthesis is compared to all other total conventional hip prostheses.

**Table 1: Revision Rates of Primary Total Conventional Hip Replacement**

Component	N Revised	N Total	Obs. Years	Revisions/100 Obs. Yrs (95% CI)
G7 Multihole	65	1256	2629	2.47 (1.91, 3.15)
Other Total Conventional Hip	19187	537317	3452113	0.56 (0.55, 0.56)
<b>TOTAL</b>	<b>19252</b>	<b>538573</b>	<b>3454742</b>	<b>0.56 (0.55, 0.57)</b>

Note: Prostheses no longer used in 2023 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 2

**Yearly Cumulative Percent Revision of Primary Total Conventional Hip Replacement**

The yearly cumulative percent revision of the G7 Multihole total conventional hip prosthesis is compared to all other total conventional hip prostheses.

**Table 2: Yearly Cumulative Percent Revision of Primary Total Conventional Hip Replacement**

CPR	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs	8 Yrs
G7 Multihole	5.2 (4.0, 6.6)	5.3 (4.1, 6.8)	5.9 (4.6, 7.6)	6.2 (4.8, 8.0)	6.2 (4.8, 8.0)			
Other Total Conventional Hip	1.7 (1.7, 1.8)	2.2 (2.1, 2.2)	2.5 (2.5, 2.5)	2.8 (2.7, 2.8)	3.1 (3.0, 3.1)	3.3 (3.3, 3.4)	3.6 (3.6, 3.7)	3.9 (3.8, 4.0)

CPR	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs
G7 Multihole								
Other Total Conventional Hip	4.2 (4.1, 4.3)	4.5 (4.5, 4.6)	4.9 (4.8, 4.9)	5.3 (5.2, 5.4)	5.7 (5.6, 5.8)	6.0 (5.9, 6.2)	6.5 (6.3, 6.6)	6.9 (6.7, 7.0)

CPR	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs
G7 Multihole							
Other Total Conventional Hip	7.3 (7.1, 7.4)	7.6 (7.5, 7.8)	8.2 (8.0, 8.4)	8.5 (8.2, 8.7)	9.0 (8.7, 9.3)	9.6 (9.2, 10.1)	10.3 (9.5, 11.2)

Note: Prostheses no longer used in 2023 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.


**FIGURE 1**

**Yearly Cumulative Percent Revision of Primary Total Conventional Hip Replacement**

The yearly cumulative percent revision of the G7 Multihole total conventional hip prosthesis is compared to all other total conventional hip prostheses. In addition, hazard ratios are reported.

Hazard ratios are reported for specific time periods during which the hazard ratio is constant. This is done to enable more specific and valid comparisons of the risk of revision over time. The pattern of variation in risk has important implications with respect to the underlying reasons for any difference.

**Figure 1: Cumulative Percent Revision of Primary Total Conventional Hip Replacement**



Number at Risk	0 Yr	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs	8 Yrs	9 Yrs	10 Yrs	11 Yrs
G7 Multihole	1256	830	603	374	184	45	8	0	0	0	0	0
Other Total Conventional Hip	537317	474841	425226	376813	332740	288937	247782	209621	174574	143301	117825	96251

Number at Risk	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs
G7 Multihole	0	0	0	0	0	0	0	0	0	0	0	0
Other Total Conventional Hip	77815	61602	47695	35979	26857	20094	14821	10048	6170	3221	1205	203

Note: Prostheses no longer used in 2023 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

**TABLE 3****Primary Diagnosis for Revised Primary Total Conventional Hip Replacement**

This table identifies the diagnosis of the primary procedure which was subsequently revised. This information is provided as there is a variation on outcome depending on the primary diagnosis. It is therefore important when considering the reasons for a higher than anticipated rate of revision that there is identification of the primary diagnosis. This information should be compared to the primary diagnosis for the revisions of all other total conventional hip prostheses.

**Table 3: Primary Diagnosis for Revised Primary Total Conventional Hip Replacement**

Primary Diagnosis	G7 Multihole		Other Total Conventional Hip	
	Number	Percent	Number	Percent
Osteoarthritis	31	47.7	15920	83.0
Fractured Neck Of Femur	9	13.8	1414	7.4
Osteonecrosis	5	7.7	854	4.5
Developmental Dysplasia	8	12.3	312	1.6
Rheumatoid Arthritis	2	3.1	206	1.1
Tumour	1	1.5	148	0.8
Failed Internal Fixation	6	9.2	145	0.8
Other Inflammatory Arthritis			106	0.6
Fracture/Dislocation	3	4.6	50	0.3
Other			17	0.1
Arthrodesis Takedown			15	0.1
<b>TOTAL</b>	<b>65</b>	<b>100.0</b>	<b>19187</b>	<b>100.0</b>

Note: Prostheses no longer used in 2023 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 4

## Reasons for Revision

This is reported in two ways: a percentage of primary procedures revised and as a percentage of all revision procedures.

**% Primaries Revised:** This shows the proportional contribution of each revision diagnosis as a percentage of the total number of primary procedures. This percentage can be used to approximate the risk of being revised for that diagnosis. Differing percentages between groups, with the same distribution of follow up time, may identify problems of concern.

**% Revisions:** The number of revisions for each diagnosis is expressed as a percentage of the total number of revisions. This shows the distribution of reasons for revision within a group but cannot be used as a comparison between groups.

Table 4: Primary Total Conventional Hip Replacement - Reason for Revision (Follow-up Limited to 6.4 Years)

Revision Diagnosis	Number	G7 Multihole		Other Total Conventional Hip		
		% Primaries Revised	% Revisions	Number	% Primaries Revised	% Revisions
Infection	22	1.8	33.8	4063	0.8	26.4
Prosthesis Dislocation/Instability	21	1.7	32.3	3842	0.7	25.0
Fracture	13	1.0	20.0	3187	0.6	20.7
Loosening	4	0.3	6.2	2659	0.5	17.3
Pain	1	0.1	1.5	284	0.1	1.8
Leg Length Discrepancy	1	0.1	1.5	283	0.1	1.8
Malposition	2	0.2	3.1	238	0.0	1.5
Implant Breakage Acetabular Insert				98	0.0	0.6
Incorrect Sizing				98	0.0	0.6
Implant Breakage Stem	1	0.1	1.5	95	0.0	0.6
Lysis				65	0.0	0.4
Implant Breakage Acetabular				49	0.0	0.3
Metal Related Pathology				43	0.0	0.3
Tumour				36	0.0	0.2
Wear Head				28	0.0	0.2
Heterotopic Bone				25	0.0	0.2
Wear Acetabular Insert				23	0.0	0.1
Implant Breakage Head				16	0.0	0.1
Wear Acetabulum				3	0.0	0.0
Osteonecrosis				2	0.0	0.0
Progression Of Disease				2	0.0	0.0
Synovitis				1	0.0	0.0
Other				246	0.0	1.6
<b>N Revision</b>	<b>65</b>	<b>5.2</b>	<b>100.0</b>	<b>15386</b>	<b>2.9</b>	<b>100.0</b>
<b>N Primary</b>	<b>1256</b>			<b>537317</b>		

Note: This table is restricted to revisions within 6.4 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2023 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

**FIGURE 2**

**Cumulative Incidence Revision Diagnosis of Primary Total Conventional Hip Replacement**

This figure details the cumulative incidence of the most common reasons for revision. The five most common reasons for revision are included as long as each of these reasons account for more than 10 procedures or at least 5% of all revisions for the G7 Multihole total conventional hip prosthesis. A comparative graph is provided of the cumulative incidence for the same reasons for revisions for all other total conventional hip prostheses.

**Figure 2: Cumulative Incidence Revision Diagnosis for Primary Total Conventional Hip Replacement**




TABLE 5

### Type of Revision Performed for Primary Total Conventional Hip Replacement

This analysis identifies the components used in the revision of the G7 Multihole total conventional hip prosthesis and compares it to the components used in the revision of all other total conventional hip prostheses.

The reason this analysis is undertaken is to identify whether there is one or more components which are being replaced that differ from the components replaced for revisions of all other total conventional hip prostheses i.e. is there a difference in the type of revision undertaken for the G7 Multihole total conventional hip prosthesis compared to all other total conventional hip prostheses.

**Table 5: Primary Total Conventional Hip Replacement - Type of Revision (Follow-up Limited to 6.4 Years)**

Type of Revision	G7 Multihole		Other Total Conventional Hip	
	Number	Percent	Number	Percent
Femoral Component	17	26.2	4790	31.1
Acetabular Component	9	13.8	2805	18.2
THR (Femoral/Acetabular)	9	13.8	1629	10.6
Cement Spacer	1	1.5	533	3.5
Removal of Prostheses			90	0.6
Reinsertion of Components			27	0.2
Bipolar Head and Femoral			6	0.0
Total Femoral			5	0.0
Saddle			1	0.0
<b>N Major</b>	<b>36</b>	<b>55.4</b>	<b>9886</b>	<b>64.3</b>
Head/Insert	26	40.0	4218	27.4
Head Only	1	1.5	860	5.6
Minor Components			244	1.6
Insert Only	2	3.1	174	1.1
Bipolar Only			2	0.0
Cement Only			1	0.0
Head/Neck			1	0.0
<b>N Minor</b>	<b>29</b>	<b>44.6</b>	<b>5500</b>	<b>35.7</b>
<b>TOTAL</b>	<b>65</b>	<b>100.0</b>	<b>15386</b>	<b>100.0</b>

Note: This table is restricted to revisions within 6.4 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2023 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

**TABLE 6****Revision Rates of G7 Multihole Primary Total Conventional Hip Replacement by Fixation**

This analysis is provided as some prostheses have more than one fixation option. Additionally there are prostheses where an alternative to the recommended approach to fixation was used e.g. a cementless prosthesis that has been cemented or vice-versa.

**Table 6: Revised Number of G7 Multihole Primary Total Conventional Hip Replacement by Fixation**

Fixation	N Revised	N Total
Cemented	2	18
Cementless	34	690
Hybrid (Femur Cemented)	29	532
Reverse Hybrid (Femur Cementless)	0	16
<b>TOTAL</b>	<b>65</b>	<b>1256</b>

**TABLE 7****Revision Rates of G7 Multihole Primary Total Conventional Hip Replacement by Bearing Surface**

This analysis is provided as some prostheses are combined with a variety of bearing surfaces. All bearing surfaces used with this prosthesis are listed.

**Table 7: Revised Number of G7 Multihole Primary Total Conventional Hip Replacement by Bearing Surface**

Bearing Surface	N Revised	N Total
Ceramic/Ceramic	0	7
Ceramic/XLPE + Antioxidant	22	478
Metal/XLPE + Antioxidant	43	760
Ceramicised Metal/XLPE + Antioxidant	0	6
Unknown	0	5
<b>TOTAL</b>	<b>65</b>	<b>1256</b>



**TABLE 8****Revision Rates of G7 Multihole Primary Total Conventional Hip Replacement by Approach**

This analysis is provided as some prostheses are used with a variety of surgical approaches. All surgical approaches used with this prosthesis are listed.

**Table 8: Revised Number of G7 Multihole Primary Total Conventional Hip Replacement by Approach**

Approach	N Revised	N Total
Anterior	5	140
Lateral	14	173
Posterior	46	931
<b>TOTAL</b>	<b>65</b>	<b>1244</b>

Note: Excludes 12 procedures with no approach recorded

TABLE 9

**Revision Rates of Primary Total Conventional Hip Replacement by State**

This enables a state by state variation to be identified for the G7 Multihole total conventional hip prosthesis and provides the comparative data for each of the states for all other total conventional hip prostheses.

The purpose of this analysis is to determine if the higher than anticipated rate of revision has widespread distribution between states. If there is widespread distribution then the reason for the higher than anticipated rate of revision is unlikely to be surgeon specific. If the prosthesis has been used in only a small number of states it is not possible to distinguish if the higher than anticipated rate of revision is related to the prosthesis, surgeon, technique or patient.

**Table 9: Revised Number of Primary Total Conventional Hip Replacement by State**

Component	State	N Revised	N Total
G7 Multihole	NSW	13	241
	VIC	13	303
	QLD	7	146
	WA	20	416
	SA	8	76
	TAS	3	50
	ACT/NT	1	24
Other Total Conventional Hip	NSW	5195	157472
	VIC	4846	140379
	QLD	3784	94450
	WA	2526	62855
	SA	1819	49327
	TAS	449	18148
	ACT/NT	568	14686
<b>TOTAL</b>		<b>19252</b>	<b>538573</b>

Note: Prostheses no longer used in 2023 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

**TABLE 10****Number of Revisions of G7 Multihole Primary Total Conventional Hip Replacement by Year of Implant**

This analysis details the number of prostheses reported each year to the Registry for the G7 Multihole total conventional hip prosthesis. It also provides the subsequent number of revisions of the primaries reported in that year.

Primary procedures performed in later years have had less follow up time therefore the number revised is expected to be less than the number revised in earlier years. For example, a primary procedure performed in 2023 has a maximum of one year to be revised, whereas a primary procedure performed in 2021 has a maximum of three years to be revised.

**Table 10: Number of Revisions of G7 Multihole Primary Total Conventional Hip Replacement by Year of Implant**

Year of Implant	Number Revised	Total Number
2017	1	15
2018	4	49
2019	13	169
2020	9	222
2021	19	243
2022	9	234
2023	10	324
<b>TOTAL</b>	<b>65</b>	<b>1256</b>

**TABLE 11****Revision Rates of G7 Multihole Primary Total Conventional Hip Replacement by Catalogue Number Range**

Many prostheses have a number of catalogue ranges. The catalogue range is specific to particular design features; more than one catalogue range usually indicates a minor difference in design in a particular G7 Multihole prosthesis.

This analysis has been undertaken to determine if the revision rate varies according to the catalogue number range.

Model	Catalogue Range	Catalogue Description	Cement	Material
<b>Acetabular</b>				
G7	110010259-110010277	G7 OSSEO TI ACETABULAR SHELL MULTI HOLE CEMENTLESS	NO	METAL

**Table 11: Revised Number of G7 Multihole Primary Total Conventional Hip Replacement by Catalogue Number Range**

Acetabular Range	N Revised	N Total
110010259-110010277	65	1256
<b>TOTAL</b>	<b>65</b>	<b>1256</b>

TABLE 12

## Revision Rates of G7 Multihole Primary Total Conventional Hip Replacement by Component

A prosthesis may be combined with multiple components. This analysis has been undertaken to determine if the revision rate varies according to the component with which it is combined.

Table 12: Revised Number of G7 Multihole Primary Total Conventional Hip Replacement by Femoral Stem Component

Femoral Stem Component	N Revised	N Total
ACTIS	0	2
AMISem H	0	4
Absolut	0	5
Accolade II	0	1
Arcos	7	53
Avenir	0	36
C-Stem AMT	0	10
CLS	1	9
CORAIL	0	38
CPCS	0	8
CPT	17	386
Echo	0	3
Evolve	1	6
Exeter V40	4	37
Generic Stem	0	1
INDIVIDUAL	0	1
KAR	0	1
M/L Taper	0	1
MS 30	3	36
Metafix	0	3
Modulus	0	1
Mutars	0	19
Optimys	0	1
Origin	0	1
Oss	2	4
Paragon	1	6
Polarstem	0	2
Quadra-C	2	6
Redapt	0	2
S-Rom	2	32
Segmental System	0	2
Short Exeter V40	1	4
Sirius	0	1
Spectron EF	0	3
Summit	0	7
Taper Fit	0	6
Taperloc	20	397
Taperloc Microplasty	2	83
VerSys	1	12
Wagner	0	6
X-Acta	0	6
ZMR	1	14
<b>TOTAL</b>	<b>65</b>	<b>1256</b>