



Australian Orthopaedic Association
National Joint Replacement Registry

2015 Annual Report – Hip and Knee Arthroplasty

ERRATUM

Amended 1 October 2015

Erratum for the **PRINTED** version of the **2015 Annual Report - Hip and Knee Arthroplasty**
The online version of the report available on the AOANJRR website has been corrected

PROSTHESES WITH HIGHER THAN ANTICIPATED RATES OF REVISION

Primary Total Conventional Hip Replacement

Page 200
Table IP7: Identified and No Longer Used
Prosthesis Combination: *Profemur TL/Dynasty

INCORRECT PRESENTATION OF DATA

Extract of Table IP7

Femoral/Acetabular	N Total	Obs. Years	Revisions/100 Obs. Yrs	Hazard Ratio, P Value
Identified and no longer used				
*Profemur L/Dynasty	193	97	6.16	Entire Period: HR=2.58 (1.16, 5.72),p=0.019

* Newly identified and no longer used

CORRECTION

Extract of Table IP7

Femoral/Acetabular	N Total	Obs. Years	Revisions/100 Obs. Yrs	Hazard Ratio, P Value
Identified and no longer used				
*Profemur TL/Dynasty	56	164	2.44	Entire Period: HR=2.77 (1.04, 7.39),p=0.041

*Newly identified and no longer used