

Optetrak-CR (cemented)/Optetrak (cemented) Total Knee Investigation

Note: This analysis compares the Optetrak-CR (ctd)/Optetrak (ctd) femoral/tibial combination with all other total knee prostheses.

This combination has been identified as having a significantly higher rate of revision. For a detailed explanation of the process used by the Registry that results in identification of prostheses that have a higher than anticipated rate of revision please refer to the Prostheses with Higher than Anticipated Rates of Revision chapter of the most recent AOANJRR Annual Report, <https://aoanjrr.sahmri.com/annual-reports-2025>.

Note: Procedures using prostheses with no recorded use in 2024 are excluded from the comparator.

TABLE 1

Revision Rate of Primary Total Knee Replacement

The revision rate of the Optetrak-CR (ctd)/Optetrak (ctd) total knee combination is compared to all other total knee prostheses.

Table 1: Revision Rates of Primary Total Knee Replacement

Component	N Revised	N Total	Obs. Years	Revisions/100 Obs. Yrs (95% CI)
Optetrak-CR (ctd)/Optetrak (ctd)	12	92	982	1.22 (0.63, 2.13)
Other Total Knee	24707	721061	4535539	0.54 (0.54, 0.55)
TOTAL	24719	721153	4536521	0.54 (0.54, 0.55)

Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 2

Yearly Cumulative Percent Revision of Primary Total Knee Replacement

The yearly cumulative percent revision of the Optetrak-CR (ctd)/Optetrak (ctd) total knee combination is compared to all other total knee prostheses.

Table 2: Yearly Cumulative Percent Revision (95% CI) of Primary Total Knee Replacement

CPR	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs	8 Yrs
Optetrak-CR (ctd)/Optetrak (ctd)	0.0 (0.0, 0.0)	3.3 (1.1, 9.8)	6.6 (3.0, 14.0)	8.9 (4.5, 17.0)	10.1 (5.4, 18.5)	10.1 (5.4, 18.5)	11.4 (6.3, 20.2)	11.4 (6.3, 20.2)
Other Total Knee	1.0 (1.0, 1.0)	1.8 (1.8, 1.9)	2.4 (2.4, 2.4)	2.8 (2.8, 2.8)	3.2 (3.1, 3.2)	3.5 (3.4, 3.5)	3.8 (3.7, 3.8)	4.1 (4.0, 4.2)

CPR	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs
Optetrak-CR (ctd)/Optetrak (ctd)	12.8 (7.3, 22.0)	12.8 (7.3, 22.0)	12.8 (7.3, 22.0)	12.8 (7.3, 22.0)	12.8 (7.3, 22.0)	12.8 (7.3, 22.0)	16.9 (9.0, 30.6)	16.9 (9.0, 30.6)
Other Total Knee	4.4 (4.3, 4.5)	4.7 (4.7, 4.8)	5.1 (5.0, 5.1)	5.4 (5.3, 5.5)	5.7 (5.6, 5.8)	6.0 (5.9, 6.1)	6.4 (6.3, 6.5)	6.8 (6.7, 7.0)


CPR	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs
Optetrak-CR (ctd)/Optetrak (ctd)	16.9 (9.0, 30.6)	16.9 (9.0, 30.6)					
Other Total Knee	7.3 (7.1, 7.4)	7.6 (7.4, 7.8)	7.9 (7.7, 8.1)	8.1 (7.9, 8.3)	8.4 (8.1, 8.6)	8.5 (8.2, 8.8)	8.6 (8.3, 8.9)

Note: Prostheses no longer used in 2024 are excluded from the comparator.

FIGURE 1**Yearly Cumulative Percent Revision of Primary Total Knee Replacement**

The yearly cumulative percent revision of the Optetrak-CR (ctd)/Optetrak (ctd) total knee combination is compared to all other total knee prostheses. In addition, hazard ratios are reported.

Hazard ratios are reported for specific time periods during which the hazard ratio is constant. This is done to enable more specific and valid comparisons of the risk of revision over time. The pattern of variation in risk has important implications with respect to the underlying reasons for any difference.

Figure 1: Cumulative Percent Revision of Primary Total Knee Replacement

Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 3**Primary Diagnosis for Revised Primary Total Knee Replacement**

This table identifies the diagnosis of the primary procedure which was subsequently revised. This information is provided as there is a variation on outcome depending on the primary diagnosis. It is therefore important when considering the reasons for a higher than anticipated rate of revision that there is identification of the primary diagnosis. This information should be compared to the primary diagnosis for the revisions of all other total knee prostheses.

Table 3: Primary Diagnosis for Revised Primary Total Knee Replacement

Primary Diagnosis	Optetrak-CR (ctd)/Optetrak (ctd)		Other Total Knee	
	Number	Percent	Number	Percent
Osteoarthritis	11	91.7	23890	96.7
Rheumatoid Arthritis			301	1.2
Tumour			192	0.8
Other Inflammatory Arthritis			158	0.6
Osteonecrosis	1	8.3	90	0.4
Fracture			50	0.2
Other			25	0.1
Chondrocalcinosis			1	0.0
TOTAL	12	100.0	24707	100.0

Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 4

Reasons for Revision

This is reported in two ways: a percentage of primary procedures revised and as a percentage of all revision procedures.

% Primaries Revised: This shows the proportional contribution of each revision diagnosis as a percentage of the total number of primary procedures. This percentage can be used to approximate the risk of being revised for that diagnosis. Differing percentages between groups, with the same distribution of follow up time, may identify problems of concern.

% Revisions: The number of revisions for each diagnosis is expressed as a percentage of the total number of revisions. This shows the distribution of reasons for revision within a group but cannot be used as a comparison between groups.

Table 4: Primary Total Knee Replacement - Reason for Revision (Follow-up Limited to 22.6 Years)

Revision Diagnosis	Optetrak-CR (ctd)/Optetrak (ctd)			Other Total Knee		
	Number	% Primaries Revised	% Revisions	Number	% Primaries Revised	% Revisions
Infection	3	3.3	25.0	7091	1.0	28.7
Loosening	7	7.6	58.3	5262	0.7	21.3
Instability	1	1.1	8.3	2537	0.4	10.3
Patella Erosion				1817	0.3	7.4
Pain				1763	0.2	7.1
Patellofemoral Pain	1	1.1	8.3	1480	0.2	6.0
Arthrofibrosis				1047	0.1	4.2
Fracture				1017	0.1	4.1
Malalignment				488	0.1	2.0
Wear Tibial Insert				322	0.0	1.3
Lysis				264	0.0	1.1
Incorrect Sizing				210	0.0	0.8
Implant Breakage Tibial Insert				204	0.0	0.8
Patella Maltracking				175	0.0	0.7
Bearing Dislocation				142	0.0	0.6
Implant Breakage Patella				132	0.0	0.5
Metal Related Pathology				101	0.0	0.4
Prosthesis Dislocation				70	0.0	0.3
Synovitis				59	0.0	0.2
Osteonecrosis				46	0.0	0.2
Implant Breakage Femoral				45	0.0	0.2
Wear Patella				43	0.0	0.2
Implant Breakage Tibial				35	0.0	0.1
Tumour				30	0.0	0.1
Heterotopic Bone				14	0.0	0.1
Progression Of Disease				8	0.0	0.0
Wear Tibial				6	0.0	0.0
Incorrect Side				1	0.0	0.0
Patella Dislocation				1	0.0	0.0
Wear Femoral				1	0.0	0.0
Other				296	0.0	1.2
N Revision	12	13.0	100.0	24707	3.4	100.0
N Primary	92			721061		

Note: This table is restricted to revisions within 22.6 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2024 are excluded from the comparator.

FIGURE 2**Cumulative Incidence Revision Diagnosis of Primary Total Knee Replacement**

This figure details the cumulative incidence of the most common reasons for revision. The five most common reasons for revision are included as long as each of these reasons account for more than 10 procedures or at least 5% of all revisions for the Optetrak-CR (ctd)/Optetrak (ctd) total knee combination. A comparative graph is provided of the cumulative incidence for the same reasons for revisions for all other total knee prostheses.


Figure 2: Cumulative Incidence Revision Diagnosis for Primary Total Knee Replacement

TABLE 5

Type of Revision Performed for Primary Total Knee Replacement

This analysis identifies the components used in the revision of the Optetrak-CR (ctd)/Optetrak (ctd) total knee combination and compares it to the components used in the revision of all other total knee prostheses.

The reason this analysis is undertaken is to identify whether there is one or more components which are being replaced that differ from the components replaced for revisions of all other total knee prostheses i.e. is there a difference in the type of revision undertaken for the Optetrak-CR (ctd)/Optetrak (ctd) total knee combination compared to all other total knee prostheses.

Table 5: Primary Total Knee Replacement - Type of Revision (Follow-up Limited to 22.6 Years)

Type of Revision	Optetrak-CR (ctd)/Optetrak (ctd)		Other Total Knee	
	Number	Percent	Number	Percent
TKR (Tibial/Femoral)	4	33.3	5912	23.9
Tibial Component	3	25.0	1803	7.3
Femoral Component			1184	4.8
Cement Spacer	1	8.3	1082	4.4
Removal of Prostheses			126	0.5
Total Femoral			22	0.1
Reinsertion of Components			7	0.0
N Major	8	66.7	10136	41.0
Insert Only	2	16.7	7667	31.0
Patella Only	2	16.7	4164	16.9
Insert/Patella			2668	10.8
Minor Components			63	0.3
Cement Only			9	0.0
N Minor	4	33.3	14571	59.0
TOTAL	12	100.0	24707	100.0

Note: This table is restricted to revisions within 22.6 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 6**Revision Rates of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Fixation**

This analysis is provided as some prostheses have more than one fixation option. Additionally there are prostheses where an alternative to the recommended approach to fixation was used e.g. a cementless prosthesis that has been cemented or vice-versa.

Table 6: Revised Number of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Fixation

Fixation	N Revised	N Total
Cemented	12	92
TOTAL	12	92

TABLE 7**Revision Rates of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Bearing Surface**

This analysis is provided as some prostheses are combined with a variety of bearing surfaces. All bearing surfaces used with this combination are listed.

Table 7: Revised Number of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Bearing Surface

Bearing Surface	N Revised	N Total
Non XLPE	12	92
TOTAL	12	92

TABLE 8**Revision Rates of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Bearing Mobility**

This analysis is provided as some prostheses are combined with a variety of bearing mobilities. All bearing mobilities used with this combination are listed.

Table 8: Revised Number of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Bearing Mobility

Bearing Mobility	N Revised	N Total
Fixed	12	92
TOTAL	12	92

TABLE 9**Revision Rates of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Stability**

This analysis is provided as some prostheses are combined with a variety of stabilities. All stabilities used with this combination are listed.

Table 9: Revised Number of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Stability

Stability	N Revised	N Total
Minimally Stabilised	12	92
TOTAL	12	92

TABLE 10

Revision Rates of Primary Total Knee Replacement by State

This enables a state by state variation to be identified for the Optetrak-CR (ctd)/Optetrak (ctd) total knee combination and provides the comparative data for each of the states for all other total knee prostheses.

The purpose of this analysis is to determine if the higher than anticipated rate of revision has widespread distribution between states. If there is widespread distribution then the reason for the higher than anticipated rate of revision is unlikely to be surgeon specific. If the prosthesis has been used in only a small number of states it is not possible to distinguish if the higher than anticipated rate of revision is related to the prosthesis, surgeon, technique or patient.

Table 10: Revised Number of Primary Total Knee Replacement by State

Component	State	N Revised	N Total
Optetrak-CR (ctd)/Optetrak (ctd)	NSW	6	43
	VIC	2	18
	QLD	4	30
	SA	0	1
Other Total Knee	NSW	6834	236304
	VIC	5457	148735
	QLD	5317	152558
	WA	3259	85158
	SA	2886	66544
	TAS	390	13242
	ACT/NT	564	18520
TOTAL		24719	721153

Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 11**Number of Revisions of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Year of Implant**

This analysis details the number of prostheses reported each year to the Registry for the Optetrak-CR (ctd)/Optetrak (ctd) total knee combination. It also provides the subsequent number of revisions of the primaries reported in that year.

Primary procedures performed in later years have had less follow up time therefore the number revised is expected to be less than the number revised in earlier years. For example, a primary procedure performed in 2024 has a maximum of one year to be revised, whereas a primary procedure performed in 2022 has a maximum of three years to be revised.

Table 11: Number of Revisions of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Year of Implant

Year of Implant	Number Revised	Total Number
2001	0	1
2002	0	6
2003	2	7
2004	1	6
2005	1	2
2006	0	9
2007	2	7
2008	0	7
2009	1	4
2011	0	5
2012	0	6
2013	0	8
2014	5	24
TOTAL	12	92

TABLE 12

Revision Rates of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Catalogue Number Range

Many prostheses have a number of catalogue ranges. The catalogue range is specific to particular design features; more than one catalogue range usually indicates a minor difference in design in a particular Optetrak-CR (ctd)/Optetrak (ctd) prosthesis.

This analysis has been undertaken to determine if the revision rate varies according to the catalogue number range.

Model	Catalogue Range	Catalogue Description	Cement
Femoral			
Optetrak-CR	2000101-2000106	CR CEMENTED FEMORAL COMPONENT	YES
Optetrak-CR	2300201-2300306	CR CEMENTED ASYMMETRIC FEMORAL COMPONENT	YES
Tibial			
Optetrak	2000421-2000612	CEMENTED FINNED TIBIAL TRAY	YES

Table 12: Revised Number of Optetrak-CR (ctd)/Optetrak (ctd) Primary Total Knee Replacement by Catalogue Number Range

Femoral Range	Tibial Range	N Revised	N Total
2000101-2000106	2000421-2000612	4	31
2300201-2300306	2000421-2000612	8	61
TOTAL		12	92