

Plasmacup Total Conventional Hip Investigation

Note: This analysis compares the Plasmacup acetabular prosthesis with all other total conventional hip prostheses.

This prosthesis has been identified as having a significantly higher rate of revision. For a detailed explanation of the process used by the Registry that results in identification of prostheses that have a higher than anticipated rate of revision please refer to the Prostheses with Higher than Anticipated Rates of Revision chapter of the most recent AOANJRR Annual Report, <https://aoanjrr.sahmri.com/annual-reports-2025>.

Note: Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator. Procedures using prostheses with no recorded use in 2024 are excluded from the comparator.

TABLE 1
Revision Rate of Primary Total Conventional Hip Replacement

The revision rate of the Plasmacup total conventional hip prosthesis is compared to all other total conventional hip prostheses.

Table 1: Revision Rates of Primary Total Conventional Hip Replacement

Component	N Revised	N Total	Obs. Years	Revisions/100 Obs. Yrs (95% CI)
Plasmacup	42	482	5331	0.79 (0.57, 1.06)
Other Total Conventional Hip	19492	552154	3552443	0.55 (0.54, 0.56)
TOTAL	19534	552636	3557774	0.55 (0.54, 0.56)

Note: Prostheses no longer used in 2024 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 2

Yearly Cumulative Percent Revision of Primary Total Conventional Hip Replacement

The yearly cumulative percent revision of the Plasmacup total conventional hip prosthesis is compared to all other total conventional hip prostheses.

Table 2: Yearly Cumulative Percent Revision (95% CI) of Primary Total Conventional Hip Replacement

CPR	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs	8 Yrs
Plasmacup	4.4 (2.9, 6.6)	5.4 (3.7, 7.8)	5.6 (3.9, 8.1)	5.6 (3.9, 8.1)	5.8 (4.1, 8.3)	6.3 (4.4, 8.8)	7.1 (5.2, 9.9)	7.4 (5.3, 10.1)
Other Total Conventional Hip	1.7 (1.7, 1.8)	2.2 (2.1, 2.2)	2.5 (2.4, 2.5)	2.8 (2.7, 2.8)	3.0 (3.0, 3.1)	3.3 (3.3, 3.4)	3.6 (3.5, 3.6)	3.9 (3.8, 3.9)

CPR	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs
Plasmacup	7.9 (5.7, 10.7)	8.1 (6.0, 11.0)	8.4 (6.2, 11.4)	8.4 (6.2, 11.4)	8.4 (6.2, 11.4)	9.0 (6.6, 12.3)	9.9 (7.1, 13.6)	
Other Total Conventional Hip	4.2 (4.1, 4.2)	4.4 (4.4, 4.5)	4.8 (4.7, 4.8)	5.2 (5.1, 5.3)	5.5 (5.4, 5.6)	5.9 (5.8, 6.0)	6.3 (6.2, 6.4)	6.7 (6.6, 6.9)

CPR	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs
Plasmacup							
Other Total Conventional Hip	7.1 (6.9, 7.2)	7.4 (7.3, 7.6)	7.9 (7.7, 8.1)	8.3 (8.0, 8.5)	8.8 (8.5, 9.1)	9.3 (9.0, 9.7)	9.9 (9.4, 10.5)

Note: Prostheses no longer used in 2024 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.


FIGURE 1

Yearly Cumulative Percent Revision of Primary Total Conventional Hip Replacement

The yearly cumulative percent revision of the Plasmacup total conventional hip prosthesis is compared to all other total conventional hip prostheses. In addition, hazard ratios are reported.

Hazard ratios are reported for specific time periods during which the hazard ratio is constant. This is done to enable more specific and valid comparisons of the risk of revision over time. The pattern of variation in risk has important implications with respect to the underlying reasons for any difference.

Figure 1: Cumulative Percent Revision of Primary Total Conventional Hip Replacement



Number at Risk	0 Yr	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs	8 Yrs	9 Yrs	10 Yrs	11 Yrs
Plasmacup	482	459	453	449	444	437	426	415	408	379	330	287
Other Total Conventional Hip	552154	487715	432817	384210	337213	295811	254946	217123	182233	150376	122151	99378

Number at Risk	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs
Plasmacup	224	159	110	64	25	21	11	6	0	0	0	0
Other Total Conventional Hip	80223	63999	49890	37936	28064	20581	15102	10871	7426	4536	2346	851

Note: Prostheses no longer used in 2024 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 3**Primary Diagnosis for Revised Primary Total Conventional Hip Replacement**

This table identifies the diagnosis of the primary procedure which was subsequently revised. This information is provided as there is a variation on outcome depending on the primary diagnosis. It is therefore important when considering the reasons for a higher than anticipated rate of revision that there is identification of the primary diagnosis. This information should be compared to the primary diagnosis for the revisions of all other total conventional hip prostheses.

Table 3: Primary Diagnosis for Revised Primary Total Conventional Hip Replacement

Primary Diagnosis	Plasmacup		Other Total Conventional Hip	
	Number	Percent	Number	Percent
Osteoarthritis	39	92.9	16174	83.0
Fractured Neck Of Femur			1436	7.4
Osteonecrosis	3	7.1	856	4.4
Developmental Dysplasia			313	1.6
Rheumatoid Arthritis			210	1.1
Failed Internal Fixation			157	0.8
Tumour			148	0.8
Other Inflammatory Arthritis			112	0.6
Fracture/Dislocation			53	0.3
Other			19	0.1
Arthrodesis Takedown			14	0.1
TOTAL	42	100.0	19492	100.0

Note: Prostheses no longer used in 2024 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 4

Reasons for Revision

This is reported in two ways: a percentage of primary procedures revised and as a percentage of all revision procedures.

% Primaries Revised: This shows the proportional contribution of each revision diagnosis as a percentage of the total number of primary procedures. This percentage can be used to approximate the risk of being revised for that diagnosis. Differing percentages between groups, with the same distribution of follow up time, may identify problems of concern.

% Revisions: The number of revisions for each diagnosis is expressed as a percentage of the total number of revisions. This shows the distribution of reasons for revision within a group but cannot be used as a comparison between groups.

Table 4: Primary Total Conventional Hip Replacement - Reason for Revision (Follow-up Limited to 19.1 Years)

Revision Diagnosis	Number	Plasmacup		Other Total Conventional Hip		
		% Primaries Revised	% Revisions	Number	% Primaries Revised	% Revisions
Infection	8	1.7	19.0	4759	0.9	24.5
Prosthesis Dislocation/Instability	4	0.8	9.5	4393	0.8	22.6
Fracture	2	0.4	4.8	4330	0.8	22.3
Loosening	23	4.8	54.8	3681	0.7	19.0
Pain				326	0.1	1.7
Leg Length Discrepancy	2	0.4	4.8	297	0.1	1.5
Malposition				267	0.0	1.4
Lysis				202	0.0	1.0
Implant Breakage Stem				199	0.0	1.0
Implant Breakage Acetabular Insert				127	0.0	0.7
Wear Acetabular Insert				101	0.0	0.5
Incorrect Sizing	1	0.2	2.4	98	0.0	0.5
Metal Related Pathology				89	0.0	0.5
Implant Breakage Acetabular				68	0.0	0.4
Wear Head				42	0.0	0.2
Tumour				40	0.0	0.2
Implant Breakage Head				30	0.0	0.2
Heterotopic Bone				27	0.0	0.1
Wear Acetabulum				10	0.0	0.1
Osteonecrosis				3	0.0	0.0
Synovitis				1	0.0	0.0
Other	2	0.4	4.8	310	0.1	1.6
N Revision	42	8.7	100.0	19400	3.5	100.0
N Primary	482			552154		

Note: This table is restricted to revisions within 19.1 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2024 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

FIGURE 2**Cumulative Incidence Revision Diagnosis of Primary Total Conventional Hip Replacement**

This figure details the cumulative incidence of the most common reasons for revision. The five most common reasons for revision are included as long as each of these reasons account for more than 10 procedures or at least 5% of all revisions for the Plasmacup total conventional hip prosthesis. A comparative graph is provided of the cumulative incidence for the same reasons for revisions for all other total conventional hip prostheses.


Figure 2: Cumulative Incidence Revision Diagnosis for Primary Total Conventional Hip Replacement

TABLE 5

Type of Revision Performed for Primary Total Conventional Hip Replacement

This analysis identifies the components used in the revision of the Plasmacup total conventional hip prosthesis and compares it to the components used in the revision of all other total conventional hip prostheses.

The reason this analysis is undertaken is to identify whether there is one or more components which are being replaced that differ from the components replaced for revisions of all other total conventional hip prostheses i.e. is there a difference in the type of revision undertaken for the Plasmacup total conventional hip prosthesis compared to all other total conventional hip prostheses.

Table 5: Primary Total Conventional Hip Replacement - Type of Revision (Follow-up Limited to 19.1 Years)

Type of Revision	Plasmacup		Other Total Conventional Hip	
	Number	Percent	Number	Percent
Femoral Component	14	33.3	6541	33.7
Acetabular Component	9	21.4	3393	17.5
THR (Femoral/Acetabular)	8	19.0	2237	11.5
Cement Spacer	3	7.1	593	3.1
Removal of Prostheses			98	0.5
Reinsertion of Components			29	0.1
Total Femoral			13	0.1
Bipolar Head and Femoral			9	0.0
N Major	34	81.0	12913	66.6
Head/Insert	3	7.1	5078	26.2
Head Only	5	11.9	922	4.8
Minor Components			305	1.6
Insert Only			179	0.9
Bipolar Only			1	0.0
Cement Only			1	0.0
Head/Neck			1	0.0
N Minor	8	19.0	6487	33.4
TOTAL	42	100.0	19400	100.0

Note: This table is restricted to revisions within 19.1 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2024 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 6**Revision Rates of Plasmacup Primary Total Conventional Hip Replacement by Fixation**

This analysis is provided as some prostheses have more than one fixation option. Additionally there are prostheses where an alternative to the recommended approach to fixation was used e.g. a cementless prosthesis that has been cemented or vice-versa.

Table 6: Revised Number of Plasmacup Primary Total Conventional Hip Replacement by Fixation

Fixation	N Revised	N Total
Cementless	35	382
Hybrid (Femur Cemented)	7	100
TOTAL	42	482

TABLE 7**Revision Rates of Plasmacup Primary Total Conventional Hip Replacement by Bearing Surface**

This analysis is provided as some prostheses are combined with a variety of bearing surfaces. All bearing surfaces used with this prosthesis are listed.

Table 7: Revised Number of Plasmacup Primary Total Conventional Hip Replacement by Bearing Surface

Bearing Surface	N Revised	N Total
Ceramic/Ceramic	37	413
Ceramic/Non XLPE	1	10
Metal/Non XLPE	4	58
Unknown	0	1
TOTAL	42	482

TABLE 8**Revision Rates of Plasmacup Primary Total Conventional Hip Replacement by Approach**

This analysis is provided as some prostheses are used with a variety of surgical approaches. All surgical approaches used with this prosthesis are listed.

Table 8: Revised Number of Plasmacup Primary Total Conventional Hip Replacement by Approach

Approach	N Revised	N Total
Anterior	3	50
Posterior	0	1
TOTAL	3	51

Note: Excludes 431 procedures with no approach recorded

TABLE 9

Revision Rates of Primary Total Conventional Hip Replacement by State

This enables a state by state variation to be identified for the Plasmacup total conventional hip prosthesis and provides the comparative data for each of the states for all other total conventional hip prostheses.

The purpose of this analysis is to determine if the higher than anticipated rate of revision has widespread distribution between states. If there is widespread distribution then the reason for the higher than anticipated rate of revision is unlikely to be surgeon specific. If the prosthesis has been used in only a small number of states it is not possible to distinguish if the higher than anticipated rate of revision is related to the prosthesis, surgeon, technique or patient.

Table 9: Revised Number of Primary Total Conventional Hip Replacement by State

Component	State	N Revised	N Total
Plasmacup	NSW	42	473
	VIC	0	4
	QLD	0	5
Other Total Conventional Hip	NSW	5288	159916
	VIC	4851	143812
	QLD	3893	100102
	WA	2492	62236
	SA	1926	51998
	TAS	448	18480
	ACT/NT	594	15610
TOTAL		19534	552636

Note: Prostheses no longer used in 2024 are excluded from the comparator. Procedures using metal/metal prostheses with head size larger than 32mm are excluded from the comparator.

TABLE 10**Number of Revisions of Plasmacup Primary Total Conventional Hip Replacement by Year of Implant**

This analysis details the number of prostheses reported each year to the Registry for the Plasmacup total conventional hip prosthesis. It also provides the subsequent number of revisions of the primaries reported in that year.

Primary procedures performed in later years have had less follow up time therefore the number revised is expected to be less than the number revised in earlier years. For example, a primary procedure performed in 2024 has a maximum of one year to be revised, whereas a primary procedure performed in 2022 has a maximum of three years to be revised.

Table 10: Number of Revisions of Plasmacup Primary Total Conventional Hip Replacement by Year of Implant

Year of Implant	Number Revised	Total Number
2005	0	10
2006	2	16
2007	2	13
2008	2	7
2009	7	54
2010	9	60
2011	6	59
2012	4	77
2013	4	70
2014	3	44
2015	2	51
2016	1	21
TOTAL	42	482

TABLE 11

Revision Rates of Plasmacup Primary Total Conventional Hip Replacement by Catalogue Number Range

Many prostheses have a number of catalogue ranges. The catalogue range is specific to particular design features; more than one catalogue range usually indicates a minor difference in design in a particular Plasmacup prosthesis.

This analysis has been undertaken to determine if the revision rate varies according to the catalogue number range.

Model	Catalogue Range	Catalogue Description	Cement	Material	Coating	Fixation
Acetabular						
Plasmacup	NC444T-NC468T	SC PLASMAPORE CAP	NO	METAL		POROUS
Plasmacup	NH140T-NH168T	MSC ACETABULAR SHELL	NO	METAL		POROUS
Plasmacup	NH340T-NH368T	NSC ACETABULAR SHELL	NO	METAL		POROUS
Plasmacup	NH544T-NH568T	DC ACETABULAR SHELL	NO	METAL		
Plasmacup	NH644D-NH654D	HA SHELL & BIOLOX DELTA CERAMIC LINER	NO	CERAMIC	HA COATED	

Table 11: Revised Number of Plasmacup Primary Total Conventional Hip Replacement by Catalogue Number Range

Acetabular Range	N Revised	N Total
NC444T-NC468T	27	289
NH140T-NH168T	1	3
NH340T-NH368T	1	13
NH544T-NH568T	0	5
NH644D-NH654D	13	172
TOTAL	42	482

TABLE 12

Revision Rates of Plasmacup Primary Total Conventional Hip Replacement by Component

A prosthesis may be combined with multiple components. This analysis has been undertaken to determine if the revision rate varies according to the component with which it is combined.

Table 12: Revised Number of Plasmacup Primary Total Conventional Hip Replacement by Femoral Stem Component

Femoral Stem Component	N Revised	N Total
Accolade I	3	27
Excia	7	92
Excia (cless)	13	164
Friendly Hip	0	1
Furlong	1	3
Hip and Go	0	25
Metha	4	80
Metha (exch neck)	14	78
S-Rom	0	7
Spectron EF	0	4
Standard C	0	1
TOTAL	42	482