

## Profix Oxinium (cementless)/Profix Mobile Total Knee Investigation

Note: This analysis compares the Profix Oxinium (cless)/Profix Mobile femoral/tibial combination with all other total knee prostheses.

This combination has been identified as having a significantly higher rate of revision. For a detailed explanation of the process used by the Registry that results in identification of prostheses that have a higher than anticipated rate of revision please refer to the Prostheses with Higher than Anticipated Rates of Revision chapter of the most recent AOANJRR Annual Report, <https://aoanjrr.sahmri.com/annual-reports-2025>.

Note: Procedures using prostheses with no recorded use in 2024 are excluded from the comparator.

**TABLE 1**

### Revision Rate of Primary Total Knee Replacement

The revision rate of the Profix Oxinium (cless)/Profix Mobile total knee combination is compared to all other total knee prostheses.

**Table 1: Revision Rates of Primary Total Knee Replacement**

Component	N Revised	N Total	Obs. Years	Revisions/100 Obs. Yrs (95% CI)
Profix Oxinium (cless)/Profix Mobile	71	158	1451	4.89 (3.82, 6.17)
Other Total Knee	24707	721061	4535539	0.54 (0.54, 0.55)
<b>TOTAL</b>	<b>24778</b>	<b>721219</b>	<b>4536990</b>	<b>0.55 (0.54, 0.55)</b>

Note: Prostheses no longer used in 2024 are excluded from the comparator.

**TABLE 2****Yearly Cumulative Percent Revision of Primary Total Knee Replacement**

The yearly cumulative percent revision of the Profix Oxinium (cless)/Profix Mobile total knee combination is compared to all other total knee prostheses.

**Table 2: Yearly Cumulative Percent Revision (95% CI) of Primary Total Knee Replacement**


CPR	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs	8 Yrs
Profix Oxinium (cless)/Profix Mobile	9.0 (5.4, 14.6)	27.0 (20.7, 34.7)	40.2 (32.9, 48.3)	41.5 (34.2, 49.7)	41.5 (34.2, 49.7)	42.9 (35.5, 51.1)	43.6 (36.2, 51.8)	43.6 (36.2, 51.8)
Other Total Knee	1.0 (1.0, 1.0)	1.8 (1.8, 1.9)	2.4 (2.4, 2.4)	2.8 (2.8, 2.8)	3.2 (3.1, 3.2)	3.5 (3.4, 3.5)	3.8 (3.7, 3.8)	4.1 (4.0, 4.2)
CPR	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs
Profix Oxinium (cless)/Profix Mobile	44.4 (36.9, 52.6)	46.0 (38.4, 54.3)	46.0 (38.4, 54.3)	46.0 (38.4, 54.3)	46.0 (38.4, 54.3)	47.0 (39.3, 55.4)	47.0 (39.3, 55.4)	47.0 (39.3, 55.4)
Other Total Knee	4.4 (4.3, 4.5)	4.7 (4.7, 4.8)	5.1 (5.0, 5.1)	5.4 (5.3, 5.5)	5.7 (5.6, 5.8)	6.0 (5.9, 6.1)	6.4 (6.3, 6.5)	6.8 (6.7, 7.0)
CPR	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs	
Profix Oxinium (cless)/Profix Mobile								
Other Total Knee	7.3 (7.1, 7.4)	7.6 (7.4, 7.8)	7.9 (7.7, 8.1)	8.1 (7.9, 8.3)	8.4 (8.1, 8.6)	8.5 (8.2, 8.8)	8.6 (8.3, 8.9)	

Note: Prostheses no longer used in 2024 are excluded from the comparator.

**FIGURE 1****Yearly Cumulative Percent Revision of Primary Total Knee Replacement**

The yearly cumulative percent revision of the Profix Oxinium (cless)/Profix Mobile total knee combination is compared to all other total knee prostheses. In addition, hazard ratios are reported.

Hazard ratios are reported for specific time periods during which the hazard ratio is constant. This is done to enable more specific and valid comparisons of the risk of revision over time. The pattern of variation in risk has important implications with respect to the underlying reasons for any difference.

**Figure 1: Cumulative Percent Revision of Primary Total Knee Replacement**

Note: Prostheses no longer used in 2024 are excluded from the comparator.

**TABLE 3****Primary Diagnosis for Revised Primary Total Knee Replacement**

This table identifies the diagnosis of the primary procedure which was subsequently revised. This information is provided as there is a variation on outcome depending on the primary diagnosis. It is therefore important when considering the reasons for a higher than anticipated rate of revision that there is identification of the primary diagnosis. This information should be compared to the primary diagnosis for the revisions of all other total knee prostheses.

**Table 3: Primary Diagnosis for Revised Primary Total Knee Replacement**

Primary Diagnosis	Profix Oxinium (cless)/Profix Mobile		Other Total Knee	
	Number	Percent	Number	Percent
Osteoarthritis	71	100.0	23890	96.7
Rheumatoid Arthritis			301	1.2
Tumour			192	0.8
Other Inflammatory Arthritis			158	0.6
Osteonecrosis			90	0.4
Fracture			50	0.2
Other			25	0.1
Chondrocalcinosis			1	0.0
<b>TOTAL</b>	<b>71</b>	<b>100.0</b>	<b>24707</b>	<b>100.0</b>

Note: Prostheses no longer used in 2024 are excluded from the comparator.

**TABLE 4****Reasons for Revision**

This is reported in two ways: a percentage of primary procedures revised and as a percentage of all revision procedures.

**% Primaries Revised:** This shows the proportional contribution of each revision diagnosis as a percentage of the total number of primary procedures. This percentage can be used to approximate the risk of being revised for that diagnosis. Differing percentages between groups, with the same distribution of follow up time, may identify problems of concern.

**% Revisions:** The number of revisions for each diagnosis is expressed as a percentage of the total number of revisions. This shows the distribution of reasons for revision within a group but cannot be used as a comparison between groups.

**Table 4: Primary Total Knee Replacement - Reason for Revision (Follow-up Limited to 22.4 Years)**


Revision Diagnosis	Profix Oxinium (class)/Profix Mobile			Other Total Knee		
	Number	% Primaries Revised	% Revisions	Number	% Primaries Revised	% Revisions
Infection	4	2.5	5.6	7091	1.0	28.7
Loosening	61	38.6	85.9	5262	0.7	21.3
Instability	2	1.3	2.8	2537	0.4	10.3
Patella Erosion				1817	0.3	7.4
Pain				1763	0.2	7.1
Patellofemoral Pain	1	0.6	1.4	1480	0.2	6.0
Arthrofibrosis	1	0.6	1.4	1047	0.1	4.2
Fracture	1	0.6	1.4	1017	0.1	4.1
Malalignment				488	0.1	2.0
Wear Tibial Insert				322	0.0	1.3
Lysis				264	0.0	1.1
Incorrect Sizing				210	0.0	0.8
Implant Breakage Tibial Insert				204	0.0	0.8
Patella Maltracking				175	0.0	0.7
Bearing Dislocation				142	0.0	0.6
Implant Breakage Patella				132	0.0	0.5
Metal Related Pathology				101	0.0	0.4
Prosthesis Dislocation	1	0.6	1.4	70	0.0	0.3
Synovitis				59	0.0	0.2
Osteonecrosis				46	0.0	0.2
Implant Breakage Femoral				45	0.0	0.2
Wear Patella				43	0.0	0.2
Implant Breakage Tibial				35	0.0	0.1
Tumour				30	0.0	0.1
Heterotopic Bone				14	0.0	0.1
Progression Of Disease				8	0.0	0.0
Wear Tibial				6	0.0	0.0
Incorrect Side				1	0.0	0.0
Patella Dislocation				1	0.0	0.0
Wear Femoral				1	0.0	0.0
Other				296	0.0	1.2
<b>N Revision</b>	<b>71</b>	<b>44.9</b>	<b>100.0</b>	<b>24707</b>	<b>3.4</b>	<b>100.0</b>
<b>N Primary</b>	<b>158</b>			<b>721061</b>		

Note: This table is restricted to revisions within 22.4 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2024 are excluded from the comparator.

**FIGURE 2****Cumulative Incidence Revision Diagnosis of Primary Total Knee Replacement**

This figure details the cumulative incidence of the most common reasons for revision. The five most common reasons for revision are included as long as each of these reasons account for more than 10 procedures or at least 5% of all revisions for the Profix Oxinium (cless)/Profix Mobile total knee combination. A comparative graph is provided of the cumulative incidence for the same reasons for revisions for all other total knee prostheses.

**Figure 2: Cumulative Incidence Revision Diagnosis for Primary Total Knee Replacement**

**TABLE 5****Type of Revision Performed for Primary Total Knee Replacement**

This analysis identifies the components used in the revision of the Profix Oxinium (cless)/Profix Mobile total knee combination and compares it to the components used in the revision of all other total knee prostheses.

The reason this analysis is undertaken is to identify whether there is one or more components which are being replaced that differ from the components replaced for revisions of all other total knee prostheses i.e. is there a difference in the type of revision undertaken for the Profix Oxinium (cless)/Profix Mobile total knee combination compared to all other total knee prostheses.

**Table 5: Primary Total Knee Replacement - Type of Revision (Follow-up Limited to 22.4 Years)**

Type of Revision	Profix Oxinium (cless)/Profix Mobile		Other Total Knee	
	Number	Percent	Number	Percent
TKR (Tibial/Femoral)	5	7.0	5912	23.9
Tibial Component			1803	7.3
Femoral Component	61	85.9	1184	4.8
Cement Spacer			1082	4.4
Removal of Prostheses			126	0.5
Total Femoral			22	0.1
Reinsertion of Components			7	0.0
<b>N Major</b>	<b>66</b>	<b>93.0</b>	<b>10136</b>	<b>41.0</b>
Insert Only	2	2.8	7667	31.0
Patella Only	1	1.4	4164	16.9
Insert/Patella	1	1.4	2668	10.8
Minor Components	1	1.4	63	0.3
Cement Only			9	0.0
<b>N Minor</b>	<b>5</b>	<b>7.0</b>	<b>14571</b>	<b>59.0</b>
<b>TOTAL</b>	<b>71</b>	<b>100.0</b>	<b>24707</b>	<b>100.0</b>

Note: This table is restricted to revisions within 22.4 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2024 are excluded from the comparator.

**TABLE 6****Revision Rates of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Fixation**

This analysis is provided as some prostheses have more than one fixation option. Additionally there are prostheses where an alternative to the recommended approach to fixation was used e.g. a cementless prosthesis that has been cemented or vice-versa.

**Table 6: Revised Number of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Fixation**

Fixation	N Revised	N Total
Cementless	55	117
Hybrid (Tibial Cemented)	16	41
<b>TOTAL</b>	<b>71</b>	<b>158</b>

**TABLE 7****Revision Rates of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Bearing Surface**

This analysis is provided as some prostheses are combined with a variety of bearing surfaces. All bearing surfaces used with this combination are listed.

**Table 7: Revised Number of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Bearing Surface**

Bearing Surface	N Revised	N Total
Non XLPE	71	158
<b>TOTAL</b>	<b>71</b>	<b>158</b>

**TABLE 8****Revision Rates of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Bearing Mobility**

This analysis is provided as some prostheses are combined with a variety of bearing mobilities. All bearing mobilities used with this combination are listed.

**Table 8: Revised Number of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Bearing Mobility**

Bearing Mobility	N Revised	N Total
Rotating	71	158
<b>TOTAL</b>	<b>71</b>	<b>158</b>

**TABLE 9****Revision Rates of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Stability**

This analysis is provided as some prostheses are combined with a variety of stabilities. All stabilities used with this combination are listed.

**Table 9: Revised Number of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Stability**

Stability	N Revised	N Total
Minimally Stabilised	71	158
<b>TOTAL</b>	<b>71</b>	<b>158</b>

**TABLE 10****Revision Rates of Primary Total Knee Replacement by State**

This enables a state by state variation to be identified for the Profix Oxinium (cless)/Profix Mobile total knee combination and provides the comparative data for each of the states for all other total knee prostheses.

The purpose of this analysis is to determine if the higher than anticipated rate of revision has widespread distribution between states. If there is widespread distribution then the reason for the higher than anticipated rate of revision is unlikely to be surgeon specific. If the prosthesis has been used in only a small number of states it is not possible to distinguish if the higher than anticipated rate of revision is related to the prosthesis, surgeon, technique or patient.

**Table 10: Revised Number of Primary Total Knee Replacement by State**

Component	State	N Revised	N Total
Profix Oxinium (cless)/Profix Mobile	NSW	44	95
	VIC	11	16
	QLD	1	7
	SA	1	1
	ACT/NT	14	39
Other Total Knee	NSW	6834	236304
	VIC	5457	148735
	QLD	5317	152558
	WA	3259	85158
	SA	2886	66544
	TAS	390	13242
<b>TOTAL</b>	ACT/NT	564	18520
		<b>24778</b>	<b>721219</b>

Note: Prostheses no longer used in 2024 are excluded from the comparator.

**TABLE 11****Number of Revisions of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Year of Implant**

This analysis details the number of prostheses reported each year to the Registry for the Profix Oxinium (cless)/Profix Mobile total knee combination. It also provides the subsequent number of revisions of the primaries reported in that year.

Primary procedures performed in later years have had less follow up time therefore the number revised is expected to be less than the number revised in earlier years. For example, a primary procedure performed in 2024 has a maximum of one year to be revised, whereas a primary procedure performed in 2022 has a maximum of three years to be revised.

**Table 11: Number of Revisions of Profix Oxinium (cless)/Profix Mobile Primary Total Knee Replacement by Year of Implant**

Year of Implant	Number Revised	Total Number
2002	28	63
2003	43	95
<b>TOTAL</b>	<b>71</b>	<b>158</b>

**TABLE 12****Revision Rates of Profix Oxinium (class)/Profix Mobile Primary Total Knee Replacement by Catalogue Number Range**

Many prostheses have a number of catalogue ranges. The catalogue range is specific to particular design features; more than one catalogue range usually indicates a minor difference in design in a particular Profix Oxinium (class)/Profix Mobile prosthesis.

This analysis has been undertaken to determine if the revision rate varies according to the catalogue number range.

Model	Catalogue Range	Catalogue Description	Cement	Coating	Fixation
<b>Femoral</b>					
Profix Oxinium	71502127-71502277	PROFIX OXINIUM MACROTEXTURED FEMORAL COMPONENT	NO		MACROTEXTURED
<b>Tibial</b>					
Profix Mobile	71544102-71544208	POROUS HA TIBIAL BASEPLATE	NO	HA COATED	POROUS
Profix Mobile	71549101-71549208	NONPOROUS TIBIAL BASEPLATE	YES		

**Table 12: Revised Number of Profix Oxinium (class)/Profix Mobile Primary Total Knee Replacement by Catalogue Number Range**

Femoral Range	Tibial Range	N Revised	N Total
71502127-71502277	71544102-71544208	56	118
	71549101-71549208	15	40
<b>TOTAL</b>		<b>71</b>	<b>158</b>