

Quadra-H Bipolar Hip Investigation

Note: This analysis compares the Quadra-H femoral stem prosthesis with all other bipolar hip prostheses.

This prosthesis has been identified as having a significantly higher rate of revision. For a detailed explanation of the process used by the Registry that results in identification of prostheses that have a higher than anticipated rate of revision please refer to the Prostheses with Higher than Anticipated Rates of Revision chapter of the most recent AOANJRR Annual Report, <https://aoanjrr.sahmri.com/annual-reports-2025>.

Note: Procedures using prostheses with no recorded use in 2024 are excluded from the comparator.

TABLE 1

Revision Rate of Primary Bipolar Hip Replacement

The revision rate of the Quadra-H bipolar hip prosthesis is compared to all other bipolar hip prostheses.

Table 1: Revision Rates of Primary Bipolar Hip Replacement

Component	N Revised	N Total	Obs. Years	Revisions/100 Obs. Yrs (95% CI)
Quadra-H	7	87	247	2.84 (1.14, 5.84)
Other Bipolar Hip	1135	38089	111864	1.01 (0.96, 1.08)
TOTAL	1142	38176	112111	1.02 (0.96, 1.08)

Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 2

Yearly Cumulative Percent Revision of Primary Bipolar Hip Replacement

The yearly cumulative percent revision of the Quadra-H bipolar hip prosthesis is compared to all other bipolar hip prostheses.

Table 2: Yearly Cumulative Percent Revision (95% CI) of Primary Bipolar Hip Replacement

CPR	1 Yr	2 Yrs	3 Yrs	4 Yrs	5 Yrs	6 Yrs	7 Yrs	8 Yrs
Quadra-H	6.2 (2.3, 15.7)	6.2 (2.3, 15.7)	10.5 (4.8, 22.3)	13.6 (6.5, 27.3)	13.6 (6.5, 27.3)	13.6 (6.5, 27.3)		
Other Bipolar Hip	2.4 (2.3, 2.6)	3.0 (2.8, 3.2)	3.4 (3.2, 3.6)	3.7 (3.5, 4.0)	4.0 (3.8, 4.3)	4.3 (4.0, 4.6)	4.5 (4.2, 4.9)	4.9 (4.5, 5.2)

CPR	9 Yrs	10 Yrs	11 Yrs	12 Yrs	13 Yrs	14 Yrs	15 Yrs	16 Yrs
Quadra-H								
Other Bipolar Hip	5.1 (4.7, 5.5)	5.4 (4.9, 5.9)	5.5 (5.0, 6.0)	5.5 (5.1, 6.1)	5.8 (5.2, 6.4)	6.1 (5.4, 6.9)	6.6 (5.8, 7.6)	7.2 (6.2, 8.5)

CPR	17 Yrs	18 Yrs	19 Yrs	20 Yrs	21 Yrs	22 Yrs	23 Yrs
Quadra-H							
Other Bipolar Hip	7.5 (6.4, 8.9)	7.5 (6.4, 8.9)	7.9 (6.6, 9.6)	7.9 (6.6, 9.6)	7.9 (6.6, 9.6)		

Note: Prostheses no longer used in 2024 are excluded from the comparator.


FIGURE 1

Yearly Cumulative Percent Revision of Primary Bipolar Hip Replacement

The yearly cumulative percent revision of the Quadra-H bipolar hip prosthesis is compared to all other bipolar hip prostheses. In addition, hazard ratios are reported.

Hazard ratios are reported for specific time periods during which the hazard ratio is constant. This is done to enable more specific and valid comparisons of the risk of revision over time. The pattern of variation in risk has important implications with respect to the underlying reasons for any difference.

Figure 1: Cumulative Percent Revision of Primary Bipolar Hip Replacement



Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 3

Primary Diagnosis for Revised Primary Bipolar Hip Replacement

This table identifies the diagnosis of the primary procedure which was subsequently revised. This information is provided as there is a variation on outcome depending on the primary diagnosis. It is therefore important when considering the reasons for a higher than anticipated rate of revision that there is identification of the primary diagnosis. This information should be compared to the primary diagnosis for the revisions of all other bipolar hip prostheses.

Table 3: Primary Diagnosis for Revised Primary Bipolar Hip Replacement

Primary Diagnosis	Quadra-H		Other Bipolar Hip	
	Number	Percent	Number	Percent
Fractured Neck Of Femur	7	100.0	1054	92.9
Osteoarthritis			32	2.8
Tumour			29	2.6
Failed Internal Fixation			10	0.9
Osteonecrosis			8	0.7
Other			2	0.2
TOTAL	7	100.0	1135	100.0

Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 4

Reasons for Revision

This is reported in two ways: a percentage of primary procedures revised and as a percentage of all revision procedures.

% Primaries Revised: This shows the proportional contribution of each revision diagnosis as a percentage of the total number of primary procedures. This percentage can be used to approximate the risk of being revised for that diagnosis. Differing percentages between groups, with the same distribution of follow up time, may identify problems of concern.

% Revisions: The number of revisions for each diagnosis is expressed as a percentage of the total number of revisions. This shows the distribution of reasons for revision within a group but cannot be used as a comparison between groups.

Table 4: Primary Bipolar Hip Replacement - Reason for Revision (Follow-up Limited to 15.2 Years)

Revision Diagnosis	Number	Quadra-H		Number	Other Bipolar Hip	
		% Primaries Revised	% Revisions		% Primaries Revised	% Revisions
Infection				354	0.9	31.3
Prosthesis Dislocation/Instability	2	2.3	28.6	298	0.8	26.3
Fracture	3	3.4	42.9	196	0.5	17.3
Loosening	1	1.1	14.3	91	0.2	8.0
Chondrolysis/Acetab. Erosion				89	0.2	7.9
Pain	1	1.1	14.3	67	0.2	5.9
Tumour				5	0.0	0.4
Lysis				4	0.0	0.4
Incorrect Sizing				3	0.0	0.3
Leg Length Discrepancy				3	0.0	0.3
Malposition				3	0.0	0.3
Heterotopic Bone				2	0.0	0.2
Implant Breakage Stem				2	0.0	0.2
Implant Breakage Acetabular				1	0.0	0.1
Implant Breakage Head				1	0.0	0.1
Metal Related Pathology				1	0.0	0.1
Osteonecrosis				1	0.0	0.1
Progression Of Disease				1	0.0	0.1
Wear Acetabulum				1	0.0	0.1
Other				8	0.0	0.7
N Revision	7	8.0	100.0	1131	3.0	100.0
N Primary	87			38089		

Note: This table is restricted to revisions within 15.2 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2024 are excluded from the comparator.

FIGURE 2

Cumulative Incidence Revision Diagnosis of Primary Bipolar Hip Replacement

This figure details the cumulative incidence of the most common reasons for revision. The five most common reasons for revision are included as long as each of these reasons account for more than 10 procedures or at least 5% of all revisions for the Quadra-H bipolar hip prosthesis. A comparative graph is provided of the cumulative incidence for the same reasons for revisions for all other bipolar hip prostheses.

Figure 2: Cumulative Incidence Revision Diagnosis for Primary Bipolar Hip Replacement




TABLE 5

Type of Revision Performed for Primary Bipolar Hip Replacement

This analysis identifies the components used in the revision of the Quadra-H bipolar hip prosthesis and compares it to the components used in the revision of all other bipolar hip prostheses.

The reason this analysis is undertaken is to identify whether there is one or more components which are being replaced that differ from the components replaced for revisions of all other bipolar hip prostheses i.e. is there a difference in the type of revision undertaken for the Quadra-H bipolar hip prosthesis compared to all other bipolar hip prostheses.

Table 5: Primary Bipolar Hip Replacement - Type of Revision (Follow-up Limited to 15.2 Years)

Type of Revision	Quadra-H		Other Bipolar Hip	
	Number	Percent	Number	Percent
Acetabular Component	3	42.9	351	31.0
THR (Femoral/Acetabular)	1	14.3	200	17.7
Bipolar Head and Femoral	2	28.6	155	13.7
Cement Spacer			40	3.5
Femoral Component	1	14.3	39	3.4
Removal of Prostheses			28	2.5
Reinsertion of Components			1	0.1
N Major	7	100.0	814	72.0
Bipolar Only			253	22.4
Head Only			39	3.4
Minor Components			24	2.1
Head/Insert			1	0.1
N Minor			317	28.0
TOTAL	7	100.0	1131	100.0

Note: This table is restricted to revisions within 15.2 years for all groups to allow a time-matched comparison of revisions.

Note: Prostheses no longer used in 2024 are excluded from the comparator.

TABLE 6**Number of Revisions of Quadra-H Primary Bipolar Hip Replacement by Year of Implant**

This analysis details the number of prostheses reported each year to the Registry for the Quadra-H bipolar hip prosthesis. It also provides the subsequent number of revisions of the primaries reported in that year.

Primary procedures performed in later years have had less follow up time therefore the number revised is expected to be less than the number revised in earlier years. For example, a primary procedure performed in 2024 has a maximum of one year to be revised, whereas a primary procedure performed in 2022 has a maximum of three years to be revised.

Table 6: Number of Revisions of Quadra-H Primary Bipolar Hip Replacement by Year of Implant

Year of Implant	Number Revised	Total Number
2009	3	11
2010	0	7
2011	1	5
2012	0	6
2013	0	4
2014	2	11
2015	0	9
2016	1	7
2017	0	4
2018	0	7
2019	0	5
2020	0	7
2021	0	1
2024	0	3
TOTAL	7	87

TABLE 7

Revision Rates of Quadra-H Primary Bipolar Hip Replacement by Component

A prosthesis may be combined with multiple components. This analysis has been undertaken to determine if the revision rate varies according to the component with which it is combined.

Table 7: Revised Number of Quadra-H Primary Bipolar Hip Replacement by Bipolar Component

Bipolar Component	N Revised	N Total
Bipolar Head (Medacta)	6	84
Multipolar Bipolar	0	1
UHR	1	2
TOTAL	7	87